



Review Article

A descriptive review on the use of skeletal anchorage in orthodontics (Part I)

N. Bhogeshwor Singh¹, Amit Nagar², Priti Shukla^{3,*}, Shraddha Gupta⁴¹Dental Officer, Imphal, Manipur, India²Dept. of Orthodontics, King George Medical University, Lucknow, Uttar Pradesh, India³Dept. of Dentistry, All India Institute of Medical Sciences, Raebareli, Uttar Pradesh, India⁴Orthodontic Practitioner, India

ARTICLE INFO

Article history:

Received 08-04-2023

Accepted 11-05-2023

Available online 03-06-2023

Keywords:

Skeletal anchorage

Branemark

Zygoma wires

ABSTRACT

Anchorage has been defined as a resistance to unwanted tooth movement. It is imperative for the treatment of both dental and skeletal malocclusions. Skeletal anchorage provides better chances to obtain absolute anchorage in orthodontic tooth movement. The topic of skeletal anchorage is vast and compiling it in one single article will be an injustice to the readers. So, we have divided this into two review articles under the headings of introduction, evolution and historical background, the means of skeletal anchorage which will be discussed in this first part followed by the design and function of screw-type orthodontic mini-implants, placement sites, surgical procedures, loading approaches, biocortical and resorbable implants and conclusion in the second part.

This is an Open Access (OA) journal, and articles are distributed under the terms of the [Creative Commons Attribution-NonCommercial-ShareAlike 4.0 License](https://creativecommons.org/licenses/by-nc-sa/4.0/), which allows others to remix, tweak, and build upon the work non-commercially, as long as appropriate credit is given and the new creations are licensed under the identical terms.

For reprints contact: reprint@ipinnovative.com

1. Introduction

Anchorage has been defined as a resistance to unwanted tooth movement. It is imperative for the treatment of both dental and skeletal malocclusions.¹ 18th century and thereafter, eminent orthodontists such as Gunnell, Desirabode, and Angle have realized upon the limitations of moving teeth against other teeth used for anchorage, introducing ideas such as the use of extraoral and occlusal anchorage.²

Newton's third law of motion must be considered during treatment planning. According to Proffit,³ during treatment planning reciprocal effects of forces that are created within the dental arches should be given due consideration and they should be controlled in order to maximize the tooth movement that is desired and to minimize the occurrence of undesirable side effects. Absolute or infinite anchorage has

been defined as no movement of the anchorage unit (zero anchorage loss) upon application of a force to move the teeth.⁴ This is possible only by means of skeletal anchorage.

2. Classification of Anchorage

Ottofy⁵ succeeded E.H. Angle in classifying anchorage as simple, stationary, reciprocal, intraoral, intermaxillary, or extraoral. Moyers⁶ expanded this classification by subcategorizing extraoral anchorage and splitting simple anchorage into single, compound, and reinforced subcategories.

Gianelly and Goldman⁷ suggested the terms maximum, moderate, and minimum based upon the extent to which the teeth of the active and reactive units move when a force is applied.

Marcotte⁸ and Burstone⁹ classified anchorage into categories—A, B, and C—based upon the anchorage unit's contribution to space closure.

* Corresponding author.

E-mail address: drprishukla22@gmail.com (P. Shukla).

Tweed¹⁰ called it anchorage preparation which was the distal tipping of posterior teeth to utilize the mechanical advantage of the tent peg causing intrusion of the molars before retracting anterior teeth.

3. Skeletal Anchorage

Skeletal Anchorage includes all the devices fixed to the bone with the goal of increasing the anchorage for orthodontic purposes.

1. Branemark implants^{11,12}
2. Retromolar implants¹¹
3. Onplants¹³
4. Zygomatic wires¹⁴
5. Ankylosed teeth¹⁵
6. Palatal implants¹⁶
7. Miniplates¹⁷ and mini screws^{18–22}

3.1. Evolution and historical background

The constant development upon the traditional orthodontic anchorage led to the evolution of Skeletal anchorage devices. Evidentially, in 1969 when Linkow²³ used a blade implant in the mandibular 1st molar region as a partial abutment before orthodontics, was the the first report concerning the use of osseointegrated implants for orthodontic purposes. Kokich²⁴ and Smalley²⁵ and Smalley and Blanco²⁶ developed protocols for accurate placement of implants for both orthodontic anchorage followed by restorative therapy.

In 1983, Creekmore and Eklund¹⁸ used a vitallium bone screw inserted in ANS to treat a patient with a deep impinging overbite. Guyman¹⁵ 1980 showed ankylosed teeth can be successfully used as anchors for palatal expansion which in 1985 was used by Kokich²⁷ to treat Apert syndrome. In 1995 Block and Hoffman¹³ designed a thin titanium alloy disk called 'onplant' being placed in fully dentate patient.

In 1996 Jurgen Glatzmaier²⁸ developed implant of biodegradable polylactide with a metal superstructure. Melsen B¹⁴ in 1988 tried zygoma ligatures as a form of maxillary anchorage, in a partially edentulous patient. In 1997 Kanomi¹⁹ introduced a miniature implant (5.0 mm x Ø1.0 mm titanium screw). In 1998, Costa²¹ presented a screw with a bracket like head to be used as orthodontic implant. Since then, multiple other implants have been introduced each presenting different designs and features.

3.2. Means of skeletal anchorage

3.2.1. Intentionally ankylosed tooth for anchorage

Guyman¹⁵ in 1980, found that ankylosed teeth can successfully be used as anchors for palatal expansion in nonhuman primates.

Advantages are²⁹

1. The tooth is biocompatible
2. Produces skeletal rather than dental movement
3. The procedure is effective with negligible risk

Disadvantage

The ankylosed tooth has a limited life expectancy before complete root resorption and exfoliation.

After Kokich et al,²⁷ Sheller²⁹ In 1991 conducted a study to develop a simple and inexpensive procedure to ankylose primary teeth, to be used for protraction.

3.2.2. Zygoma wires

Melsen B¹⁴ in 1998 introduced the Zygomatic ligature which is an inexpensive and simple method of anchorage for mechanics required in maxillary incisors. He suggests that, the best bone quality in a partially edentulous patient is found in the region of the zygomatic arch and the infrazygomatic crest which can be used for anchorage purposes.

Advantages:

1. No special equipment needed
2. Inexpensive material
3. Immediate use post insertion for anchorage
4. Lesser treatment duration
5. Removal is quick and easy

3.2.3. Conventional dental implants

The number of orthodontic adult patients is increasing who are often partially edentulous. For using conventional dental implants as anchors in such case firstly orthodontic treatment is completed and the implant is used later as abutments for fixed restorations.

3.2.4. Palatal implants and onplants

Palatal implants and onplants are miniature, osseointegrated devices, but because they are removed after orthodontic treatment, they qualify as Temporary Anchorage Devices (TADs). Used in fully dentate arches which cannot accommodate implants in the alveolar process and also do not hamper space closure. Consequently, implants have to be placed in other locations.³⁰ For the maxilla, both the midsagittal³¹ and paramedian³² regions of the hard palate have been proposed for easy accessibility and excellent peri-implant conditions being covered by attached mucosa.

Palatal implants are osseointegrated and can be connected by a transpalatal arch (TPA), thereby offering absolute orthodontic anchorage.³³

Advantages of palatal implants include –

1. Easy to use.
2. Increased stability.
3. No requirement of patient cooperation.
4. Increased aesthetics.

Palatal implants compensate for small length by having a machined or modified surface (SLA meaning sand blasted, large grit, acid).

Currently palatal implants are available in two systems-

1. The Straumann Orthosystem.
2. The flange fixture.

The Straumann Orthosystem was developed by Wehrbein.¹⁶ The flange fixture is primarily used for anchoring facial prostheses but due to its small length it is also been used in the palate.³³

Uses of palatal implants-

1. For creating or closing maxillary spaces
2. For maxillary mesialization or distalization
3. For correcting asymmetries involving midline and intercuspation
4. In cases of partial edentulism of posterior region such that the anterior teeth can be altered three dimensionally.
5. Adult maxillary expansion

3.3. Resorbable palatal implants

In 1996 Jurgen Glatzmaier²⁸ developed implant of biodegradable polylactide with a metal superstructure which resorb without a foreign body reaction and prevents the need for surgical intervention during removal.

The volume of bone available for placement of palatal implants in the hard palate is determined by using Lateral cephalograms and cone beam computed tomography presurgically.³⁴

3.4. Orthodontic Mechanics with palatal implants

The loading of the implants can be done either directly or indirectly depending on the clinical situation and the treatment plan. Both the standard and the chair-side procedure can be used for connecting the teeth to be stabilized with the palatal implant.^{35,36}

By direct loading we mean that the forces that are required for the achievable tooth movement are directly introduced onto the implant. The translatory movements which are the resultant are achieved with a constant force with the help of NiTi springs or E-chains as seen with levers.³⁴

3.5. Removal procedure

For removal of the orthodontic implant, osseointegration needs to be broken which is done by minimally invasive counterclockwise rotation using a force of up to 55 Ncm.³⁴

3.6. Complications of palatal implants

1. Loss of implant caused by peri-implantitis and implant loosening.

2. Perforation of nasal floor or maxillary sinus.

During the initial healing period, slightly mobile implants may gain stability within 6 weeks.

Use of Chlorhexidine digluconate rinses thrice daily and mechanical cleaning with a soft toothbrush may control loosening of the implant.

3.7. Palatal onplants

In 1995 Block and Hoffman,¹³ designed 'onplant' which is a thin titanium alloy disk for use as TAD. The onplant surface that lies against the bone is textured and coated with a 75 μm thick layer of hydroxyapatite while the one facing the soft tissue is smooth titanium alloy with an internal threaded hole at its center for abutment placement. The placement of the onplant is similar to that of the palatal implant. It has a healing period of approximately 12 weeks after which orthodontic forces are loaded on the abutment.

Block and Hoffman¹³ in their animal study showed that it provides sufficient anchorage to successfully move and anchor teeth.

It was Xiang Chen³⁶ in 2007 who investigated the biomechanical properties and the degree of osseointegration of onplants during various healing periods in an animal model.

Feldmann³⁷ in their investigation evaluated the orthodontic anchorage capacity of 4 anchorage systems during all phases of maxillary extractions cases.

Bantleon³⁸ (2002) reported a 92% success rate for osseointegration in a subjective report of 40 Orthosystem palatal implants.

3.7.1. Miniplates

In 1999, Unemori et al¹⁷ reported the use of miniplates for posterior intrusion in anterior open bite cases. In comparison to other TADs miniplates are advantageous as they do not interfere with tooth movement and the use of multiple screws provide more secure anchorage which is especially beneficial in patients with extremely thin cortical bone.

Different systems using miniplates as anchors –

1. Skeletal Anchorage Systems (SAS by Sugawara³⁹ in 2000
2. Zygoma Anchorage System (ZAS by De Clerck⁴⁰ in 2002

4. Skeletal Anchorage Systems

Sugawara³⁹ devised skeletal anchorage system (SAS) utilizing titanium miniplates and monocortical screws that are temporarily fixed in the jaws for absolute anchorage. It offers a nonsurgical, as well as a nonextraction treatment approach for maxillary or mandibular protrusion, and/or anterior crowding in adult patients, retreatment cases, and patients with complex orthodontic problems.

5. Appliance Design

The SAS comprises of bone plates and fixation screws³⁹ which are made of biocompatible pure titanium which is suitable for osseointegration. The anchor plate consists of the head, the arm, and the body. The head is exposed intraorally and positioned outside of the dentition so that it does not interfere with tooth movement. The arm is transmucosal while the body is positioned subperiosteally and is available in three different configurations—

1. The T-plate,
2. The Y-plate, and
3. The I-plate

The site of placement for the Y-plate is the zygomatic buttress to intrude or distalize upper molars, the I-plate is the anterior ridge of the piriform opening for intrusion of upper anterior teeth or protraction of upper molars and the T-plate is the mandibular body to intrude, protract, or distalize lower molars, or at the anterior border of the ascending ramus to extrude impacted molars. Costa recommend the use of computerized tomographs along with traditionally used panoramic and periapical x-rays

5.1. Complications

Most patients who undergo SAS placement show mild to moderate facial swelling for several days after surgery. Infection onsite has been reported in about 10% of patients. Mild infections can be controlled by use of antiseptic mouthwash and careful brushing techniques. In more severe cases, antibiotics are required. Other potential complications include plate fracture and mucosal dehiscence around the plate.

5.1.1. Advantages of SAS

1. Bio-compatible.
2. Most rigid available skeletal anchorage.
3. Noninterference in tooth movement due to site of location.
4. No need for patient compliances.
5. Predictable treatment results.
6. Decreased need of extraction and surgery.

Quite similar to SAS, De Clerck⁴⁰ in 2002 devised Zygoma Anchorage System (ZAS) using titanium miniplate. The zygomatic anchorage system is one of the safe anchorage methods. Because of the location and solid bone structure, the inferior border of the zygomaticomaxillary buttress, between the first and second molars, was chosen as the implant site, near the center of resistance of the first permanent molar.

5.2. Main indications

1. Enmasse distalization.

2. Mesialization of posteriors.
3. Intrusion of a single tooth or a group of a teeth;
4. Orthopedic intermaxillary traction.

Removal of ZAS is under local anesthesia through a small vertical incision in the gingival covering the miniplate. A special screwdriver, that fits into the pentagonal outer holes of the screw heads is used.

5.3. Advantages of ZAS

1. The ZAS uses three miniscrews, thus increasing total anchorage.
2. Immediate loading is possible
3. The point of application of the orthodontic forces is brought down to the level of the furcation of the upper first molar roots.
4. Does not interfere with tooth movement.
5. The vertical slot with the locking screw makes it possible to attach an auxiliary wire, which can move the point of force application some distance from the anchor.

6. Miniscrew - Implants as Anchors in Orthodontics

The treatment using micro-implant is independent upon patient compliance, make treatment time shorter, and can achieve good result. The Micro-implants could provide absolute anchorage and have revolutionized the orthodontic treatment options. These are also called as Temporary Anchorage Devices.

7. Temporary Anchorage Devices

A temporary anchorage device (TAD) is a device that is temporarily fixed to bone for the purpose of enhancing orthodontic anchorage either by supporting the teeth of the reactive unit or by obviate the need for the reactive unit altogether, and which is subsequently removed after use. They can be located transosteally, subperiosteally, or endosteally; and they can be fixed to bone either mechanically (cortically stabilized) or biochemically (osseo-integrated). TADs into orthodontic treatment made possible infinite anchorage, showing no movement (zero anchorage loss) as a consequence of reaction forces.²

Synonyms used to describe devices of skeletal anchorage.

1. Mini-implants,
2. Microscrew implant,
3. Micro-implant,
4. Minidental implant,
5. Screw-type implant,
6. Intraoral anchorage systems,
7. Temporary anchorage devices

8. Conclusion

Skeletal orthodontic anchorage has majorly changed the possibilities and paradigms in orthodontic treatment. It obviates the need for significant patient compliance, particularly with regard to extraoral appliances, which allows more predictable treatment results. The second part to follow will continue with the design and function of screw-type orthodontic mini-implants, placement sites, surgical procedures, loading approaches, biocortical and resorbable implants and conclusion.

9. Conflicts of Interests

The authors have no financial interests or conflicts of interests.

10. Source of Funding

None.

References

- Weinberger BW. The history of orthodontia - Part 6. *Int J Orthod.* 1916;2:103–17.
- Daskalogiannakis J. Glossary of orthodontic terms. Leipzig: Quintessence Publishing Co; 2000. p. 5–6.
- Proffit WR. Contemporary orthodontics. 3rd ed. St. Louis: Mosby; 2000. p. 147–95.
- Melsen B, Garbo D. Treating the “impossible case” with the use of the Aarhus Anchorage System. *Orthod.* 2004;1:13–20.
- Ottofy L. Standard Dental Dictionary. Chicago: Laird and Lee, Inc; 1923. p. 49.
- Moyers R. Handbook of Orthodontics for the Student and General Practitioner. Chicago: Year Book Medical Publishers Inc; 1973. p. 440–2.
- Gianelly A, Goldman H. Biologic Basis of Orthodontics. Philadelphia: Lea and Febiger; 1971. p. 224.
- Marcotte M. Biomechanics in Orthodontics. Toronto: BD Decker; 1990. p. 221–54.
- Burstone CJ. En masse space closure. In: Modern Edgewise Mechanics and the Segmented Arch Technique. Glendore: Ornicor Corp; 1995. p. 50–60.
- Tweed C. Clinical Orthodontics. St Louis, CV Mosby Co; 1966. p. 7–12.
- Roberts WE, Helm FR, Marshall KJ, and RKG. Rigid endosseous implants for orthodontic and orthopedic anchorage. *Angle Orthod.* 1989;59(4):247–56. doi:10.1043/0003-3219(1989)059<0247:REIFOA>2.0.CO;2.
- Sherman A. Bone reaction to orthodontic forces on vitreous carbon dental implants. *Am J Orthod.* 1978;74(1):79–87. doi:10.1016/0002-9416(78)90047-7.
- Block MS, Hoffman DR. A new device for absolute anchorage for orthodontics. *Am J Orthod Dentofacial Orthop.* 1995;107(3):251–8. doi:10.1016/s0889-5406(95)70140-0.
- Melsen B, Petersen JK, Costa A. Zygoma ligatures: an alternative form of maxillary anchorage. *J Clin Orthod.* 1998;32(3):154–8.
- Guyman GW, Kokich VG, Oswald RJ. Ankylosed teeth as abutments for palatal expansion in the rhesus monkeys. *Am J Orthod.* 1980;77(5):486–99. doi:10.1016/0002-9416(80)90128-1.
- Wehrbein H, Glatzmaier J, Mundwiler U, Diedrich P. The orthosystem: A new implant system for orthodontic anchorage in the palate. *J Orofac Orthop.* 1996;57(3):142–53. doi:10.1007/BF02191878.
- Umemori M, Sugawara J, Nagasaka H, Kawamura H. Skeletal anchorage system for open-bite correction. *Am J Orthod Dentofacial Orthop.* 1999;115(2):166–74. doi:10.1016/S0889-5406(99)70345-8.
- Creekmore TD, Eklund MK. The possibility of skeletal anchorage. *J Clin Orthod.* 1983;17(4):266–9.
- Kanomi R, Higle LB. A study of orthodontic anchorage possibilities in basal bone. *J Clin Orthod Oral Surg.* 1997;31(8):763–70. doi:10.1016/0096-6347(45)90025-1.
- Gainsforth BL, Higley LB. A study of orthodontic anchorage possibilities in basal bone. *Am J Orthod Oral Surg.* 1945;31(8):406–16. doi:10.1016/0096-6347(45)90025-1.
- Costa A, Raffling M, Millstone B, B. Miniscrews as orthodontic anchorage: A preliminary report. *Int J Adult Orthod Orthog Surg.* 1998;13(3):201–9.
- Maino BG, Bednar J, Pagin P, Mura P. The spider screw for skeletal anchorage. *J Clin Orthod.* 2003;37(2):90–7.
- Linkow LI. The endosseous blade implant and its use in orthodontics. *Int J Orthod.* 1969;7(4):149–54.
- Kokich VG. Managing complex orthodontic problems: the use of implants for anchorage. *Semin Orthod.* 1996;2(2):153–60. doi:10.1016/s1073-8746(96)80050-8.
- Smalley WM. Implants for tooth movement: determining implant location and orientation. *J Esthet Dent.* 1995;7(2):62–72. doi:10.1111/j.1708-8240.1995.tb00564.x.
- Smalley WM, Blanco A. Implants for tooth movement: a fabrication and placement technique for provisional restorations. *J Esthet Dent.* 1995;7(4):150–4. doi:10.1111/j.1708-8240.1995.tb00571.x.
- Kokich VG, Shapiro PA, Oswald R, Koskinen ML, Clarren SK. Ankylosed teeth as abutments for maxillary protraction: A case report. *Am J Orthod.* 1985;88(4):303–7. doi:10.1016/0002-9416(85)90129-0.
- Glatzmaier J, Wehrbein H, Oiedrich P. Biodegradable implants for orthodontic anchorage. A preliminary biomechanical study. *Eur J Orthod.* 1996;18(5):465–9. doi:10.1093/ejo/18.5.465.
- Brunski J, Jr AM, Pollack SR, Korostoff E, Trachtenberg DI. The influence of functional use of endosseous dental implants on the tissue-implant interface: Histologic aspects. *J Dent Res.* 1953;58(10):1953–69. doi:10.1177/00220345790580100201.
- Wehrbein H, Diedrich P. Endosseous titanium implants during and after orthodontic load: an experimental study in dog. *Clin Oral Implant Res.* 1993;4(2):76–82. doi:10.1034/j.1600-0501.1993.040203.x.
- Wehrbein H, Merz BR, Diedrich P. Palatal bone support for orthodontic implant anchorage: a clinical and radiological study. *Eur J Orthod.* 1999;21(1):65–70. doi:10.1093/ejo/21.1.65.
- Bernhart T, Vollgruber A, Gahleitner A, Dörtbudak O, Haas R. Alternative to the median region of the palate for placement of an orthodontic implant. *Clin Oral Implants Res.* 2000;11(6):595–601. doi:10.1034/j.1600-0501.2000.011006595.x.
- Bernhart T, Freudenthaler J, Dörtbudak O, Bantleon HP, Watzek G. Short epithetic implants for orthodontic anchorage in the paramedian region of the palate: a clinical study. *Clin Oral Implant Res.* 2001;12(6):624–31. doi:10.1034/j.1600-0501.2001.120611.x.
- Crismani AG, Bernhart T, Bantleon H, Cope JB. Palatal implants: The Straumann Orthosystem. *Semin Orthod.* 2005;11(1):16–23. doi:10.1053/j.sodo.2004.11.004.
- Crismani AG, Bernhart T, Baier C. Chair-side procedure for connecting transpalatal arches with palatal implants. *Eur J Orthod.* 2002;24(4):337–42. doi:10.1093/ejo/24.4.337.
- Chen X, Chen G, He H, Peng C, Zhang T. Peter Ngan Osseointegration and biomechanical properties of the onplant system. *Am J Orthod Dentofacial Orthop.* 2007;132:278.e1–6.
- Feldmann I, Bondemark L. Anchorage capacity of osseointegrated and conventional anchorage systems: A randomized controlled trial. *Am J Orthod Dentofacial Orthop.* 2008;133(3):339.19–28. doi:10.1016/j.ajodo.2007.08.014.
- Bantleon H, Bernhart T, Crismani AG, Zachrisson BU. Stable orthodontic anchorage with palatal osseointegrated implants. *World J Orthod.* 2002;3:109–16.
- Sugawara J, Nishimura M. Minibone implants: The skeletal anchorage system. *Semin Orthod.* 2005;11(1):47–56. doi:10.1053/j.sodo.2004.11.008.

40. De Clerck H, Geerinckx V, Siciliano S. The zygoma anchorage system. *J Clin Orthod*. 2002;36(8):455–9.

Priti Shukla, Assistant Professor  <https://orcid.org/0000-0002-7251-7776>

Author biography

Shraddha Gupta, Orthodontic Practitioner

N. Bhogeshwor Singh, Dental Officer

Amit Nagar, Professor  <https://orcid.org/0000-0001-5899-1482>

Cite this article: Singh NB, Nagar A, Shukla P, Gupta S. A descriptive review on the use of skeletal anchorage in orthodontics (Part I). *IP Indian J Orthod Dentofacial Res* 2023;9(2):77-82.